

BREATH INSTRUMENT
CALIBRATION TEST RECORD

Dept. NEW HARTFORD POLICE DEPARTMENT

Inst. No. 127606

DATE	SOLUTION LOT NO.	TEST RESULTS			TEMP.	OPERATOR	SIMULATOR EQUILIBRA
		1	2	3			
10/20/19						RFI Defect - error - reset Dynamometer Power	next test marker DCJS with full mic
10/16/19	19180	.10	.10	.10	34.01	Kamusta	MP4198
10/23/19	19180	.10	.10	.10	34.01	"	"
10/30/19	19180	.09	.07	.10	34.0	"	"
11/2/19	19180					new solution	"
11/2/19	19180	.10	.10	.10	34.0	Kamusta	"
11/8/19	19180	.10	.10	.10	34.01	"	"
11/15/19	19180	.10	.10	.10	34.01	"	"
11/22/19	19180	.10	.10	.10	34.01	"	"
11/29/19	19250					new solution	"
11/29/19	19250	.10	.10	.10	34.0	Kamusta	"
12/6/19	19250	.10	.10	.10	34.0	"	"
12/12/19	19250	.10	.10	.10	34.0	"	"

BREATH INSTRUMENT
CALIBRATION TEST RECORD

Dept. NEW HARTFORD POLICE DEPARTMENT

Inst. No. 127606

DATE	SOLUTION LOT NO.	TEST RESULTS			TEMP.	OPERATOR	SIMULATOR EQUILIBRA
		1	2	3			
6/1/19	19090					New Solution	MP4/199
6/1/19	19090	.09	.10	.10	34.0	Kowalski	" "
6/5/19	19090					New Solution - All Benz	" "
6/5/19	19090	.10	.10	.10	34.0	Kowalski	" "
6/12/19	19090	.09	.10	.10	34.0	" "	" "
6/20/19	19090	.10	.10	.10	34.0	" "	" "
6/26/19	19090	.10	.10	.10	34.0	" "	" "
7/1/19						New Solution	" "
7/1/19	19090	.10	.10	.10	34.0	Kowalski	" "
7/8/19	19090	.10	.10	.10	34.0	" "	" "
7/15/19	19090	.10	.10	.10	34.0	" "	" "
7/22/19	19090	.10	.10	.10	34.0	Fontaine	" "
7/29/19	19090	.10	.10	.10	34.0	Kowalski	" "
7/31/19	19130					New Solution	" "
7/31/19	19130	.09	.10	.10	34.0	Kowalski	" "
8/7/19	19130	.10	.10	.10	34.0	" "	" "
8/14/19	19130	.10	.10	.10	34.0	" "	" "
8/21/19	19130	.10	.10	.10	34.0	" "	" "
8/28/19	19130	.10	.10	.10	34.0	" "	" "
9/1/19	19130					New Solution	" "
9/3/19	19130	.10	.10	.10	34.0	Kowalski	MP4/199
9/10/19	19130	.10	.10	.10	34.0	" "	" "
9/18/19	19130	.10	.10	.10	34.0	" "	" "
9/25/19	19130	.10	.10	.10	34.0	" "	" "
9/30/19	19130					New Solution	" "
9/30/19	19130	.10	.10	.10	34.0	Kowalski	" "
10/9/19	19130	.10	.10	.10	34.0	" "	" "

BREATH INSTRUMENT
CALIBRATION TEST RECORD

Dept. NEW HARTFORD POLICE DEPARTMENT

Inst. No. 127606

DATE	SOLUTION LOT NO.	TEST RESULTS			TEMP.	OPERATOR	SIMULATOR EQUILIBRA
		1	2	3			
1/23/19	18310	.09	.09	.09	34.0	Kowalski	MP4199
1/30/19	18260					new Solution	" "
1/30/19	18360	.10	.10	.10	34.0	Kowalski	" "
2/6/19	18360	.10	.10	.10	34.0	" "	" "
2/12/19	18360	.10	.10	.10	34.0	Fontaine	" "
2/20/19	18360	.10	.10	.10	34.0	Kowalski	" "
2/27/19	18360	.10	.09	.09	34.0	" "	" "
3/1/19	18300					new Solution	" "
3/1/19	18300	.10	.10	.10	34.0	Kowalski	" "
3/8/19	18300	.10	.10	.10	34.0	" "	" "
3/15/19	18300	.10	.10	.10	34.0	" "	" "
3/21/19	18360	.09	.09	.10	34.0	" "	" "
3/28/19	18360	.09	.09	.09	34.0	" "	" "
3/31/19	19030					new Solution	" "
3/31/19	19030	.10	.10	.10	34.0	Kowalski	" "
4/4/19	19030	.10	.10	.10	34.0	" "	" "
4/11/19	19030	.10	.09	.10	34.0	" "	" "
4/18/19	19030	.09	.10	.10	34.0	" "	" "
4/24/19	19030	.09	.10	.09	34.0	" "	" "
5/1/19	19030					new Solution	" "
5/1/19	19030	.09	.10	.10	34.0	Kowalski	" "
5/8/19	19030	.09	.10	.10	34.0	" "	" "
5/15/19	19030	.09	.10	.10	34.0	" "	" "
5/21/19	19030	.09	.09	.09	34.0	" "	" "
5/24/19	19030	.09	.09	.09	34.0	" "	" "
5/31/19	19030	.09	.09	.09	34.0	" "	" "

out of service - Ethernet connection issue
sent to
By Police