

SUPERVISOR TEST

NEW HARTFORD TOWN
POLICE DEPT
DATAMASTER dmt: 127606

Date: 02/12/2020

OPERATOR NAME:
MICHAEL T KOWALSKI
Permit #: 31077
Operator Agency/Dept: NEW HARTFORD TOWN PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:34
INTERNAL STANDARD	VERIFIED	07:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:35
BLANK TEST	0.00	07:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:36
BLANK TEST	0.00	07:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:37
BLANK TEST	0.00	07:38

