

SUPERVISOR TEST

NEW HARTFORD TOWN
POLICE DEPT
DATAMASTER dmt: 127606

Date: 02/19/2020

OPERATOR NAME:
MICHAEL T KOWALSKI
Permit #: 31077
Operator Agency/Dept: NEW HARTFORD TOWN PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:30
INTERNAL STANDARD	VERIFIED	07:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:30
BLANK TEST	0.00	07:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:31
BLANK TEST	0.00	07:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:33
BLANK TEST	0.00	07:33

