

SOLUTION CHANGE

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 06/02/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:50
INTERNAL STANDARD	VERIFIED	12:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:51
BLANK TEST	0.00	12:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:52
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:54

