

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 05/20/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:10
INTERNAL STANDARD	VERIFIED	13:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:10
BLANK TEST	0.00	13:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:12
BLANK TEST	0.00	13:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:13
BLANK TEST	0.00	13:14

