

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 11/20/2019

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:20
INTERNAL STANDARD	VERIFIED	16:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:20
BLANK TEST	0.00	16:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:21
BLANK TEST	0.00	16:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:23
BLANK TEST	0.00	16:23

