

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 10/16/2019

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:25
INTERNAL STANDARD	VERIFIED	17:25
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	17:25
BLANK TEST	0.00	17:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:27
BLANK TEST	0.00	17:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:28
BLANK TEST	0.00	17:29

