

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197




Date: 12/26/2019

OPERATOR NAME:
MICHAEL P GIBBONS
Permit #: 45972
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:03
INTERNAL STANDARD	VERIFIED	03:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:03
BLANK TEST	0.00	03:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:05
BLANK TEST	0.00	03:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:06
BLANK TEST	0.00	03:07

WFO Weekly
Datamaster



SUPERVISOR TEST WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Karcic

Date: 01/30/2020

OPERATOR NAME:
SAHID KARCIC
Permit #: 46876
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:04
INTERNAL STANDARD	VERIFIED	02:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:05
BLANK TEST	0.00	02:06
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:06
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:07
BLANK TEST	0.00	02:08

