

SOLUTION CHANGE

-----  
ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197  
-----




Date: 02/01/2020

OPERATOR NAME:  
MICHAEL P GIBBONS  
Permit #: 45972  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:16
INTERNAL STANDARD	VERIFIED	19:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:17
BLANK TEST	0.00	19:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:18
BLANK TEST	0.00	19:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:19
BLANK TEST	0.00	19:20

*WFO  
Weekly*



SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197




Date: 02/12/2020

OPERATOR NAME:  
MICHAEL P GIBBONS  
Permit #: 45972  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:42
INTERNAL STANDARD	VERIFIED	00:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:43
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:44
BLANK TEST	0.00	00:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:45
BLANK TEST	0.00	00:46

WFO  
Weekly Test



SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 02/12/2020

OPERATOR NAME:  
BRANDON J RUTLEDGE  
Permit #: 43762  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:19
INTERNAL STANDARD	VERIFIED	22:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:19
BLANK TEST	0.00	22:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:20
BLANK TEST	0.00	22:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:21
BLANK TEST	0.00	22:22

WFO  
WEEKLY

