

BFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906



Date: 05/28/2020

OPERATOR NAME:
EDWARD F CONSTANTINE
Permit #: 45980
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:35
INTERNAL STANDARD	VERIFIED	03:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:36
BLANK TEST	0.00	03:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:37
BLANK TEST	0.00	03:38

RFI Detected



BFC

SUPERVISOR TEST




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 05/28/2020

OPERATOR NAME:
EDWARD F CONSTANTINE
Permit #: 45980
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:42
INTERNAL STANDARD	VERIFIED	03:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:42
BLANK TEST	0.00	03:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:44
BLANK TEST	0.00	03:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:45
BLANK TEST	0.00	03:46



SOLUTION CHANGE

BFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/02/2020

OPERATOR NAME:
BRETT A GRABELDINGER
Permit #: 46896
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:45
INTERNAL STANDARD	VERIFIED	03:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:45
BLANK TEST	0.00	03:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:47
BLANK TEST	0.00	03:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:48
BLANK TEST	0.00	03:49



BFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/03/2020

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:13
INTERNAL STANDARD	VERIFIED	18:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:13
BLANK TEST	0.00	18:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:15
BLANK TEST	0.00	18:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:16
BLANK TEST	0.00	18:17

