

SUPERVISOR TEST

LES

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606



LATIC

Date: 03/12/2020

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:43
INTERNAL STANDARD	VERIFIED	05:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:44
BLANK TEST	0.00	05:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:45
BLANK TEST	0.00	05:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:46
BLANK TEST	0.00	05:47



SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/18/2020

OPERATOR NAME:
BRETT A GRABELDINGER
Permit #: 46896
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:30
INTERNAL STANDARD	VERIFIED	17:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:30
BLANK TEST	0.00	17:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:31
BLANK TEST	0.00	17:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:33
BLANK TEST	0.00	17:34

