

SUPERVISOR TEST *CFD*

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 04/15/2020

OPERATOR NAME:  
BRETT A GRABELDINGER  
Permit #: 46896  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:28
INTERNAL STANDARD	VERIFIED	22:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:29
BLANK TEST	0.00	22:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:30
BLANK TEST	0.00	22:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:31
BLANK TEST	0.00	22:32

