

SUPERVISOR TEST




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/01/2020

OPERATOR NAME:
LAUREN C MARLEAU
Permit #: 45975
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:07
INTERNAL STANDARD	VERIFIED	14:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:07
BLANK TEST	0.00	14:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:08
BLANK TEST	0.00	14:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:10
BLANK TEST	0.00	14:11



SOLUTION CHANGE

BFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/01/2020

OPERATOR NAME:
BRETT A GRABELDINGER
Permit #: 46896
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:30
INTERNAL STANDARD	VERIFIED	20:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:30
BLANK TEST	0.00	20:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:31
BLANK TEST	0.00	20:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:32
BLANK TEST	0.00	20:33



TECHNICIAN TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906





Date: 04/07/2020

Nominal = 0.100 g/210L

BLANK TEST	0.000	14:33
INTERNAL STANDARD	VERIFIED	14:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.102	14:33
BLANK TEST	0.000	14:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.102	14:34
BLANK TEST	0.000	14:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.102	14:36
BLANK TEST	0.000	14:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.102	14:37
BLANK TEST	0.000	14:38

Average = 0.1020
Std Dev = 0.0000

BFO



SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 04/08/2020

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:21
INTERNAL STANDARD	VERIFIED	20:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:22
BLANK TEST	0.00	20:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:23
BLANK TEST	0.00	20:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:24
BLANK TEST	0.00	20:25

BFO

