

BFO

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 03/25/2020

OPERATOR NAME:  
LEONARD F DILLENBECK  
Permit #: 47621  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:10
INTERNAL STANDARD	VERIFIED	20:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:11
BLANK TEST	0.00	20:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:13
BLANK TEST	0.00	20:14

