

SOLUTION CHANGE

SHERRILL CITY
POLICE DEPT
DATAMASTER dmt: 115108

Date: 03/01/2020

OPERATOR NAME:
SHAWN R PLATT
Permit #: 30289
Operator Agency/Dept: SHERRILL CITY PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:41
INTERNAL STANDARD	VERIFIED	08:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:41
BLANK TEST	0.00	08:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:42
BLANK TEST	0.00	08:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:44
BLANK TEST	0.00	08:44

