

SUPERVISOR TEST

-----  
SHERRILL CITY  
POLICE DEPT  
DATAMASTER dmt: 115108  
-----

Date: 11/20/2019

OPERATOR NAME:  
ROBERT J DRAKE  
Permit #: 27298  
Operator Agency/Dept: SHERRILL CITY PD  
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:03
INTERNAL STANDARD	VERIFIED	07:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:04
BLANK TEST	0.00	07:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:05
BLANK TEST	0.00	07:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:06
BLANK TEST	0.00	07:07

