

SUPERVISOR TEST

-----  
SHERRILL CITY  
POLICE DEPT  
DATAMASTER dmt: 115108  
-----

Date: 03/25/2020

OPERATOR NAME:  
ROBERT J DRAKE  
Permit #: 27298  
Operator Agency/Dept: SHERRILL CITY PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:22
INTERNAL STANDARD	VERIFIED	07:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:23
BLANK TEST	0.00	07:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:25
BLANK TEST	0.00	07:26

