

```

<?xml version="1.0" standalone="yes"?>
<TRACS_IMPORT>
  <ALCOTEST_VALUES>
    <SERIAL>ARAH-0027</SERIAL>
    <IP_ADRESS>10.81.202.29</IP_ADRESS>
    <DATE>01/26/2020</DATE>
    <TIME>02:42AM</TIME>
  </ALCOTEST_VALUES>
  <CALCHECK_1>
    <CAL_MESSUNG_2_MESSUNG_ERROR>0</CAL_MESSUNG_2_MESSUNG_ERROR>
    <CAL_MESSUNG_2_IR_AAK_NULLWERT_1>0.00</CAL_MESSUNG_2
_IR_AAK_NULLWERT_1>
    <CAL_MESSUNG_2_IR_AAK_MESSWERT>0.09</CAL_MESSUNG_2
_IR_AAK_MESSWERT>
    <CAL_MESSUNG_2_EC_AAK_MESSWERT>0.10</CAL_MESSUNG_2
_EC_AAK_MESSWERT>
    <CAL_MESSUNG_2_IR_AAK_NULLWERT_2>0.00</CAL_MESSUNG_2
_IR_AAK_NULLWERT_2>
    <ID_BLANK_TIME_BEFORE_2>02:38:31AM 01/26/2020
</ID_BLANK_TIME_BEFORE_2>
    <ID_TIME_START_BLOW_2>--:--:-- --/--/-----
</ID_TIME_START_BLOW_2>
    <ID_TIME_MEAS_2>02:39:08AM 01/26/2020</ID_TIME_MEAS_2>
    <ID_BLANK_TIME_AFTER_2>02:39:43AM 01/26/2020
</ID_BLANK_TIME_AFTER_2>
  </CALCHECK_1>
  <SUBJECT_TEST>
    <PROB_MESSUNG_2_MESSUNG_ERROR>8</PROB_MESSUNG_2
_MESSUNG_ERROR>
    <PROB_MESSUNG_2_IR_AAK_NULLWERT_1>0.00</PROB_MESSUNG_2
_IR_AAK_NULLWERT_1>
    <PROB_MESSUNG_2_IR_AAK_MESSWERT>--</PROB_MESSUNG_2
_IR_AAK_MESSWERT>
    <PROB_MESSUNG_2_EC_AAK_MESSWERT>--</PROB_MESSUNG_2
_EC_AAK_MESSWERT>
    <PROB_MESSUNG_2_IR_AAK_NULLWERT_2>0.00</PROB_MESSUNG_2
_IR_AAK_NULLWERT_2>
    <PROB_MESSUNG_2_BLAS_FEHLVERSUCHE>0</PROB_MESSUNG_2
_BLAS_FEHLVERSUCHE>
    <PROB_MESSUNG_2_BLAS_VOLUMEN>3.2</PROB_MESSUNG_2
_BLAS_VOLUMEN>
    <PROB_MESSUNG_2_BLAS_ZEIT>8.6</PROB_MESSUNG_2_BLAS_ZEIT>
    <ID_BLANK_TIME_BEFORE_3>02:40:51AM 01/26/2020
</ID_BLANK_TIME_BEFORE_3>
    <ID_TIME_START_BLOW_3>02:41:39AM 01/26/2020
</ID_TIME_START_BLOW_3>
    <ID_TIME_END_BLOW_3>02:41:47AM 01/26/2020</ID_TIME_END_BLOW_
3>
    <ID_TIME_MEAS_3>02:41:56AM 01/26/2020</ID_TIME_MEAS_3>
    <ID_BLANK_TIME_AFTER_3>02:43:11AM 01/26/2020
</ID_BLANK_TIME_AFTER_3>
  </SUBJECT_TEST>

```

```

<CALCHECK_2>
  <CAL_MESSUNG_3_MESSUNG_ERROR>-</CAL_MESSUNG_3_MESSUNG_ERROR>
  <CAL_MESSUNG_3_IR_AAK_NULLWERT_1>-</CAL_MESSUNG_3
_IR_AAK_NULLWERT_1>
  <CAL_MESSUNG_3_IR_AAK_MESSWERT>-</CAL_MESSUNG_3
_IR_AAK_MESSWERT>
  <CAL_MESSUNG_3_EC_AAK_MESSWERT>-</CAL_MESSUNG_3
_EC_AAK_MESSWERT>
  <CAL_MESSUNG_3_IR_AAK_NULLWERT_2>-</CAL_MESSUNG_3
_IR_AAK_NULLWERT_2>
  <ID_BLANK_TIME_BEFORE_4>--:--:-- --/--/-----
</ID_BLANK_TIME_BEFORE_4>
  <ID_TIME_START_BLOW_4>--:--:-- --/--/-----
</ID_TIME_START_BLOW_4>
  <ID_TIME_MEAS_4>--:--:-- --/--/-----</ID_TIME_MEAS_4>
  <ID_BLANK_TIME_AFTER_4>--:--:-- --/--/-----
</ID_BLANK_TIME_AFTER_4>
</CALCHECK_2>
<GENERAL_DATA>
  <ID_SERIAL_NO>ARAH-0027</ID_SERIAL_NO>
  <ID_SW_VERSION> 8321288 1.12.2</ID_SW_VERSION>
  <ID_START_TIME>02:32:36AM 01/26/2020</ID_START_TIME>
  <ID_END_TIME>02:43:11AM 01/26/2020</ID_END_TIME>
  <ID_LAST_CAL_DATE>02/12/2019</ID_LAST_CAL_DATE>
  <ID_AAK_ENDRESULT>INTERFERENT DETECTED</ID_AAK_ENDRESULT>
  <ID_TEST_COUNTER>732</ID_TEST_COUNTER>
  <ID_MEAS_PROC>1</ID_MEAS_PROC>
  <ID_CALIBRATION_COUNTER>70</ID_CALIBRATION_COUNTER>
  <ID_ERROR_PROTOKOLL>8</ID_ERROR_PROTOKOLL>
  <ID_TEST_GAS_TYPE>2</ID_TEST_GAS_TYPE>
  <ID_TEST_GAS_INLET>1</ID_TEST_GAS_INLET>
  <ID_TEST_GAS_CONC>25.00</ID_TEST_GAS_CONC>
  <ID_TEST_CFG_SENSOR>0</ID_TEST_CFG_SENSOR>
  <ID_DATA_INPUT_ON_OFF>1</ID_DATA_INPUT_ON_OFF>
  <ID_IR_AVERAGE>21684</ID_IR_AVERAGE>
  <ID_EC_SENSOR_OFFSET>109</ID_EC_SENSOR_OFFSET>
  <ID_HOSE_TEMP>46.9</ID_HOSE_TEMP>
  <ID_IR_SENSOR_TEMP>45.3</ID_IR_SENSOR_TEMP>
  <ID_INTERNAL_DEVICE_TEMP>36.7</ID_INTERNAL_DEVICE_TEMP>
  <ID_12V_VOLTAGE>12.1</ID_12V_VOLTAGE>
  <ID_3V_BACKUP_VOLTAGE>3.1</ID_3V_BACKUP_VOLTAGE>
  <ID_MS_SW_CRC>0x6581</ID_MS_SW_CRC>
  <ID_PRESSURE_DRY_GAS>0</ID_PRESSURE_DRY_GAS>
  <ID_DRY_GAS_LOT_NO>874951 CYL 62</ID_DRY_GAS_LOT_NO>
  <ID_DRY_GAS_EXP_DATA>12/05/2020</ID_DRY_GAS_EXP_DATA>
</GENERAL_DATA>
<INPUT_VALUES>
  <UF_CASE_NUMBER>AL20-0130</UF_CASE_NUMBER>
  <ZERO_TOLERANCE_CASE>NO</ZERO_TOLERANCE_CASE>
  <TROOP>T</TROOP>
  <STATION>ALBANY</STATION>
  <ARREST_DATE>01/26/2020</ARREST_DATE>

```

<ARREST_TIME>02:09AM</ARREST_TIME>
<ARRESTING_MEM_LAST>LISCHAK</ARRESTING_MEM_LAST>
<ARRESTING_MEM_FIRST>MARK</ARRESTING_MEM_FIRST>
<ARRESTING_MEM_MI>A</ARRESTING_MEM_MI>
<ARRESTING_MEM_SHIELD>2168</ARRESTING_MEM_SHIELD>
<SUBJECT_LAST_NAME>MARTINEZ OLIVO</SUBJECT_LAST_NAME>
<SUBJECT_FIRST_NAME>ROBERTO</SUBJECT_FIRST_NAME>
<SUBJECT_MI>--</SUBJECT_MI>
<SUBJECT_DOB>10/03/1986</SUBJECT_DOB>
<SUBJECT_AGE>33</SUBJECT_AGE>
<SUBJECT_GENDER>MALE</SUBJECT_GENDER>
<OPERATOR_LAST_NAME>LISCHAK</OPERATOR_LAST_NAME>
<OPERATOR_FIRST_NAME>MARK</OPERATOR_FIRST_NAME>
<OPERATOR_MI>A</OPERATOR_MI>
<OPERATOR_SHIELD_NO>2168</OPERATOR_SHIELD_NO>
<OPERATOR_PERMIT_NO>43296</OPERATOR_PERMIT_NO>
<OPERATOR_PERMIT_EXP>01/09/2021</OPERATOR_PERMIT_EXP>
<SUBJECT_OBSERVED_20_MINUTES>YES</SUBJECT_OBSERVED_20
_MINUTES>
 <TEST_20_MINUTES_PAST_ARREST>YES</TEST_20
_MINUTES_PAST_ARREST>
 <REFUSAL_MIRANDA_READ>YES</REFUSAL_MIRANDA_READ>
</INPUT_VALUES>
</TRACS_IMPORT>