



**State
Police**

**Forensic
Investigation
Center**

ANDREW M. CUOMO
Governor

KEITH M. CORLETT
Superintendent

DR. RAY A. WICKENHEISER
Director

0.10% SIMULATOR SOLUTION RECORD

Report Date: December 11, 2019
Manufacturer: Guth Laboratories, Inc.
Lot Number: 19350
Expiration Date: 5/25/2020
Analysis Results: 122 mg Ethyl Alcohol / 100 mL of solution

I, Carol S Miller, tested simulator solution lot number 19350 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

I, Carol S Miller, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:

Reviewed by:

Digitally signed under ESRA
by Carol S Miller on
2019.12.11 10:22:24 -05'00'

Digitally signed under
ESRA by Carrie A Kirkton on
2019.12.13 15:06:55 -05'00'

Carol S Miller

Forensic Scientist III

Breath Analysis Technical Supervisor/Operator #43419

Carrie A Kirkton

Supervisor of Forensic Services

Breath Analysis Technical Supervisor