

SOLUTION CHANGE

NEW YORK MILLS  
POLICE DEPT.  
DATAMASTER dmt: 500475

Date: 01/31/2020

OPERATOR NAME:  
ROBERT C FRANKLAND  
Permit #: 40674  
Operator Agency/Dept: NY MILLS PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 17:18 |
| INTERNAL STANDARD       | VERIFIED | 17:19 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 17:19 |
| BLANK TEST              | 0.00     | 17:20 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 17:20 |
| BLANK TEST              | 0.00     | 17:21 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 17:21 |
| BLANK TEST              | 0.00     | 17:22 |

