

SOLUTION CHANGE

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 11/30/2019

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:32
INTERNAL STANDARD	VERIFIED	23:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:33
BLANK TEST	0.00	23:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:34
BLANK TEST	0.00	23:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:35
BLANK TEST	0.00	23:36

