

SUPERVISOR TEST

NEW YORK MILLS  
POLICE DEPT.  
DATAMASTER dmt: 500475

Date: 12/18/2019

OPERATOR NAME:  
ROBERT C FRANKLAND  
Permit #: 40674  
Operator Agency/Dept: NY MILLS PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:06
INTERNAL STANDARD	VERIFIED	11:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:06
BLANK TEST	0.00	11:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:08
BLANK TEST	0.00	11:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:09
BLANK TEST	0.00	11:10

