

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 03/18/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:34
INTERNAL STANDARD	VERIFIED	15:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:35
BLANK TEST	0.00	15:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:36
BLANK TEST	0.00	15:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:37
BLANK TEST	0.00	15:38

