

SUPERVISOR TEST

NEW YORK MILLS  
POLICE DEPT.  
DATAMASTER dmt: 500475

Date: 04/22/2020

OPERATOR NAME:  
ROBERT C FRANKLAND  
Permit #: 40674  
Operator Agency/Dept: NY MILLS PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:55
INTERNAL STANDARD	VERIFIED	14:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:56
BLANK TEST	0.00	14:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:57
BLANK TEST	0.00	14:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:58
BLANK TEST	0.00	14:59

