

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 03/25/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:08
INTERNAL STANDARD	VERIFIED	20:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:09
BLANK TEST	0.00	20:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:10
BLANK TEST	0.00	20:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	20:11
BLANK TEST	0.00	20:12

