

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: Oneida County
Sheriff's Office

Instrument Serial #: ~~100197~~
100197

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
12/6/18	18310	34.02	34.02	34.02	MP4192	.10	.10	.10	Grabbe/Dinger	No
12/13/18	18310	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
12/14/18	18310	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
12/26/18	18310	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
01/02/19	18310	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
01/09/19	18310	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
01/16/19	18310	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
01/23/19	18310	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
1/30/19	18360	34.02	34.02	34.02	MP4192	.10	.10	.10	Grabbe/Dinger	yes
2/6/19	18360	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
2/13/19	18360	34.02	34.02	34.02	MP4192	.10	.10	.10	Stanek	No
2/20/19	18360	34.02	34.02	34.01	MP4192	.10	.10	.10	Gibbons	No

* - Please enter a √ to denote simulator solution change taking place

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: _____

Instrument Serial #: _____

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
2/28/19	18360	34.02	34.02	34.02	MP4192	.10	.10	.10	Stanek	No
03/06/19	18360	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
3/13/19	18360	34.02	34.02	34.02	MP4192	.10	.10	.10	Stanek	No
3/20/19	18360	34.03	34.03	34.03	MP4192	.10	.10	.10	Stanek	No
3/27/19	18360	34.03	34.03	34.03	MP4192	.10	.10	.10	Stanek	No
04/02/19	19030	34.03	34.03	34.03	MP4192	.10	.10	.10	Gibbons	Yes
4/3/19	19030	34.03	34.03	34.03	MP4192	.10	.10	.10	Stanek	No
4/11/19	19030	34.03	34.03	34.03	MP4192	.10	.10	.10	Stanek	No
4/18/19	19030	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
4/24/19	19030	34.02	34.02	34.02	MP4192	.10	.10	.10	Stanek	No
5/1/19	19030	34.02	34.03	34.03	MP4192	.10	.10	.10	Grabbeinger	No-Bottle Change
5/8/19	19030	34.02	34.02	34.02	MP4192	.10	.10	.10	Stanek	No

* - Please enter a √ to denote simulator solution change taking place

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: OCSO - WFO

Instrument Serial #: 100197

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
5/16/19	19030	34.03	34.03	34.03	MP4192	.10	.10	.10	Gibbons	No
5/23/19	19030	34.02	34.02	34.02	MP4192	.10	.10	.10	Stancok	No
5/30/19	19090	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	Yes
6/8/19	19090	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
6/12/19	19090	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
6/19/19	19090	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
6/27	19090	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
7/4	19090	34.02	34.02	34.02	MP4192	.10	.10	.10	LATIC	Yes
7/10	19090	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
7/18	19090	34.03	34.03	34.02	MP4192	.10	.10	.10	Gibbons	No
7/24	19090	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
8-1	19130	34.02	34.02	34.02	MP4192	.10	.10	.10	Rutledge	Yes

* - Please enter a √ to denote simulator solution change taking place

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: OCSSO - WFO

Instrument Serial #: 100197

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
8/7/19	19130	34.02 34.02	34.03 34.03	34.03 34.03	M24192	.10	.10	.10	Gibbons	No
8/14/19	19130	34.02	34.02	34.02	M24192	.10	.10	.10	RUTLEDGE	No
8/22/19	19130	34.02	34.02	34.02	M24192	.10	.10	.10	Gibbons	No
8/28/19	19130	34.02	34.02	34.02	M24192	.10	.10	.10	Bowers	No
9/1/19	19130	34.02	34.02	34.02	M24192	.10	.10	.10	Gibbons	No - Bottle Yes change
9/4/19	19130	34.02	34.02	34.02	M24192	.10	.10	.10	Gibbons	No
9/11/19	19130	34.02	34.02	34.02	M24192	.10	.10	.10	Bowers	No
9/18/19	19130	34.02	34.02	34.02	M24192	.10	.10	.10	Gibbons	No
9/26/19	19130	34.02	34.02	34.02	M24192	.10	.10	.10	Bowers	No
9/29/19	19180	34.02	34.02	34.02	M24192	.10	.10	.10	Grabbe/river	Yes
10/16/19	19180	34.02	34.02	34.03	M24192	.10	.10	.10	Grabbe/river	No
10/23/19	19180	34.02	34.03	34.02	M24192	.10	.10	.10	Gibbons	No

* - Please enter a √ to denote simulator solution change taking place

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: _____

Instrument Serial #: _____

OC 50 - WFO

10019.7

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
11-7-19	19180	34.02 34.02	34.02	34.02	MP4192	.10	.10	.10	LATIC	NO
11/13/19	19180 53215	34.03	34.02	34.03	MP4192	.10	.10	.10	Dillenbeck	NO
11/26/19	19180	34.02	34.03	34.03	MP4192	.10	.10	.10	LATIC	NO
11/27/19	19180	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
12/3/19	19250	34.02	34.03	34.03	MP4192	.10	.10	.10	Grabbe, ror	YES
12/6/19	19250	34.02	34.02	34.02	MP4192	.10	.10	.10	LATIC	NO
12/11/19	19250	34.03	34.03	34.02	MP4192	.10	.10	.10	Gibbons	No
12/18/19	19250	34.02	34.02	34.02	MP4192	.10	.10	.10	LATIC	NO
12/26/19	19250	34.03	34.03	34.03	MP4192	.10	.10	.10	Gibbons	NO
01/05/20	19250	34.02	34.03	34.03	MP4192	.10	.10	.10	Gibbons	NO

* - Please enter a √ to denote simulator solution change taking place