

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197




Date: 04/01/2020

OPERATOR NAME:
MICHAEL P GIBBONS
Permit #: 45972
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:28
INTERNAL STANDARD	VERIFIED	22:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:28
BLANK TEST	0.00	22:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:29
BLANK TEST	0.00	22:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:30
BLANK TEST	0.00	22:31

WFO Datamaster
Weekly Test



SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 04/08/2020

OPERATOR NAME:
BRANDON J RUTLEDGE
Permit #: 43762
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:06
INTERNAL STANDARD	VERIFIED	21:06
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:07
BLANK TEST	0.00	21:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:08
BLANK TEST	0.00	21:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:09
BLANK TEST	0.00	21:10

WFO

