

SUPERVISOR TEST

wfo




ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 01/15/2020

OPERATOR NAME:  
EMRAH LATIC  
Permit #: 46908  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 19250

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 23:35 |
| INTERNAL STANDARD       | VERIFIED | 23:35 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.10     | 23:35 |
| BLANK TEST              | 0.00     | 23:36 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.10     | 23:36 |
| BLANK TEST              | 0.00     | 23:37 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 23:37 |
| BLANK TEST              | 0.00     | 23:38 |



SUPERVISOR TEST

WFO

LATIC




ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 01/15/2020

OPERATOR NAME:  
EMRAH LATIC  
Permit #: 46908  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 19250

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 23:35 |
| INTERNAL STANDARD       | VERIFIED | 23:35 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.10     | 23:35 |
| BLANK TEST              | 0.00     | 23:36 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.10     | 23:36 |
| BLANK TEST              | 0.00     | 23:37 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 23:37 |
| BLANK TEST              | 0.00     | 23:38 |



SUPERVISOR TEST

WFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 01/22/2020

OPERATOR NAME:  
BRETT A GRABELDINGER  
Permit #: 46896  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 19250

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 19:52 |
| INTERNAL STANDARD       | VERIFIED | 19:52 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.10     | 19:52 |
| BLANK TEST              | 0.00     | 19:53 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.10     | 19:53 |
| BLANK TEST              | 0.00     | 19:54 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.10     | 19:54 |
| BLANK TEST              | 0.00     | 19:55 |

