

BFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Dillenbeck

Date: 05/20/2020

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:44
INTERNAL STANDARD	VERIFIED	23:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:45
BLANK TEST	0.00	23:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:46
BLANK TEST	0.00	23:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:47
BLANK TEST	0.00	23:48

