

SUPERVISOR TEST

LATIC

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606




LEB

Date: 05/07/2020

OPERATOR NAME:  
EMRAH LATIC  
Permit #: 46908  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:00
INTERNAL STANDARD	VERIFIED	02:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:00
BLANK TEST	0.00	02:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:02
BLANK TEST	0.00	02:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:03
BLANK TEST	0.00	02:04



SUPERVISOR TEST

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

LATIC

Date: 05/12/2020

OPERATOR NAME:  
EMRAH LATIC  
Permit #: 46908  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:56
INTERNAL STANDARD	VERIFIED	17:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:57
BLANK TEST	0.00	17:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:58
BLANK TEST	0.00	17:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:59
BLANK TEST	0.00	18:00

