

Oneida County Sheriff's Department  
Breath Test Instrument  
Camden Field Office

900A - 9304314-001  
DATAMASTER-900162

123906

Date	Operator	Defendant	Chemical Test	Purge Analysis	Instrument Number	Arresting Officer	
7/4	BAKER	Weekly	.09X3	34.00-34.01	123906	19090	
7/10/19	Grabeldinger	weekly	.09X3	34.01X3	123906	19090	
7/19/19	LATIC	Weekly	.09X3	34.00x2 34.01x1	123906	19090	
7/24/19	Bowers	Weekly	.09X3	34.00 34.01x2	123906	19090	
8/1/19	Constantine	Weekly	.09X3	34.00x1 34.01x1 34.02x1	123906	<del>19090</del> 19130	
8/7/19	Grabeldinger	weekly	.09X3	34.01X3	123906	19130	
8/14/19	Bowers	Weekly	.09X3	34.00x2 34.01x1	123906	19130	
8/21/19	OSIEKA	Weekly	.09X3	34.01	123906	19130	
8/28/19	Bowers	Weekly	.09X3	34.00x1 34.01x2	123906	19130	
9/1/19	Grabeldinger	Bottle Change	.09-10	34.01X3	123906	19130	
9/4/19	Bowers	Weekly	.09X3	34.00x2 34.01x1	123906	19130	
9/17/19	Bowers	James E. Wheelock	.27	34.00	123906	Hoffman	
9/11/19	Bowers	Weekly	.09X3	34.00x3	123906	19130	
9/19/19	Grabeldinger	Weekly	.09X3	34.00-34.01	123906	19130	
9/26/19	LATIC	Weekly	.09X3	34.00 x2	34.01 x1	123906	19130
10/9/19	LATIC	Weekly - solution change	.10x3	34.01 34.01x2	123906	19150	
10/16/19	Grabeldinger	weekly	.10X3	34.01X3	123906	19180	
10/19/19	LATIC	John A. McCormack	.14	34-01	123906	19180	
11/7/19	Bowers	Weekly	.10X3	34.00x1 34.01x2	123906	19180	
11/14/19	Grabeldinger	weekly	.10X3	34.00- 34.01	123906	19180	
11-20-19	LATIC	Weekly	.10x3	34.01x2 34.02x1	123906	19180	
11/27/19	Grabeldinger	weekly	.10X3	34.00- 34.01	123906	19180	
12/2/19	Grabeldinger	Solution Change	.10X3	34.00- 34.01	123906	19250	

