

SOLUTION CHANGE

SHERRILL CITY
 POLICE DEPT
 DATAMASTER dmt: 115108

Date: 01/31/2020

OPERATOR NAME:
 SHAWN R PLATT
 Permit #: 30289
 Operator Agency/Dept: SHERRILL CITY PD
 Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:40
INTERNAL STANDARD	VERIFIED	18:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:40
BLANK TEST	0.00	18:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:41
BLANK TEST	0.00	18:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:43
BLANK TEST	0.00	18:44

