

SUPERVISOR TEST

SHERRILL CITY  
POLICE DEPT  
DATAMASTER dmt: 115108

Date: 01/29/2020

OPERATOR NAME:  
SHAWN R PLATT  
Permit #: 30289  
Operator Agency/Dept: SHERRILL CITY PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:12
INTERNAL STANDARD	VERIFIED	08:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:12
BLANK TEST	0.00	08:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:14
BLANK TEST	0.00	08:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16

