

SUPERVISOR TEST

-----  
SHERRILL CITY  
POLICE DEPT  
DATAMASTER dmt: 115108  
-----

Date: 05/06/2020

OPERATOR NAME:  
ROBERT J DRAKE  
Permit #: 27298  
Operator Agency/Dept: SHERRILL CITY PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:09
INTERNAL STANDARD	VERIFIED	07:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:10
BLANK TEST	0.00	07:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:11
BLANK TEST	0.00	07:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:12
BLANK TEST	0.00	07:13

