

**Preliminary Breath Test Device  
Monthly Verification Log**

Department   NYPD  

Device Serial #   02052756  

Test Date	Sim. Solution Lot #	Simulator Temperature			Sim. Serial #	Test Results			Officer Performing Verification	Sim. Solution Change *
		1	2	3		1	2	3		

10/27/18	18230	33.9	33.9	33.9	3701	.09	.09	.09	Komisi	/
12/7/18	18310	33.9	33.9	33.9	3701	.09	.09	.09	" "	/
1/30/19	18360	33.9	33.9	33.9	3701	.10	.10	.10	" "	/
2/29/19	18360	34.0	33.9	33.9	" "	.09	.09	.09	" "	/
3/31/19	18360	33.9	33.9	33.9	" "	.09	.09	.09	" "	
4/29/19	19010	33.9	33.9	33.9	" "	.09	.09	.09	" "	/
5/31/19	19050	33.9	33.9	33.9	" "	.09	.09	.09	" "	
6/28/19	19090	34.0	33.9	33.9	" "	.09	.09	.09	" "	/
7/25/19	19450	33.9	33.9	33.9	" "	.09	.09	.09	" "	/
8/30/19	19130	34.0	34.0	33.9	" "	.09	.09	.09	" "	/
8/30/19	19130	33.9	33.9	34.0	" "	.09	.09	.09	" "	/
10/30/19	19180	33.9	33.9	33.9	" "	.09	.09	.09	" "	/
11/29/19	19180	34.0	33.9	34.0	" "	.09	.09	.09	" "	✗
12/29/19	19250	33.9	33.9	33.9	" "	.09	.09	.09	" "	/
2/7/20	19350	33.9	33.9	33.9	" "	.10	.10	.10	" "	/

\* - Please enter a ✓ to denote a simulator solution lot number change

Version 8/12/2015

**Preliminary Breath Test Device  
Monthly Verification Log**

Department WHPD Device Serial # 007462

Test Date	Sim. Solution Lot #	Simulator Temperature			Sim. Serial #	Test Results			Officer Performing Verification	Sim. Solution Change *
		1	2	3		1	2	3		

2/7/20	19350	33.9	33.9	33.9	3705	1.09	.09	.09	Kewensco	✓

\* - Please enter a ✓ to denote a simulator solution lot number change  
Version 8/12/2015

**Preliminary Breath Test Device  
Monthly Verification Log**

Department NYPD

Device Serial # 097315

Test Date	Sim. Solution Lot #	Simulator Temperature			Sim. Serial #	Test Results			Officer Performing Verification	Sim. Solution Change *
		1	2	3		1	2	3		

2/27/19	18360	34.0	34.0	33.9	3201	.09	.09	.09	Kamster	—
3/31/19	18360	34.0	34.0	34.0	" "	.09	.09	.09	" "	—
4/30/19	19030	33.9	33.9	33.9	" "	.09	.09	.09	" "	✓
6/7/19	19090	33.8	32.8	33.8	" "	.09	.09	.09	" "	—
6/28/19	19090	33.8	33.9	33.9	" "	.09	.09	.09	" "	—
7/18/19	19090	33.8	33.8	33.9	" "	.09	.09	.09	" "	—
8/20/19	19130	33.9	33.9	34.0	" "	.09	.09	.09	" "	—
9/30/19	19130	34.0	33.9	33.9	" "	.09	.09	.09	" "	—
10/30/19	19180	33.9	33.9	33.9	" "	.09	.09	.09	" "	—
11/29/19	19180	34.0	33.9	33.9	" "	.09	.09	.09	" "	—
12/29/19	19250	33.9	33.9	33.9	" "	.09	.09	.09	" "	—
2/7/20	19350	33.9	34.0	33.9	" "	.09	.09	.09	" "	✓

\* - Please enter a ✓ to denote a simulator solution lot number change

Version 8/12/2015

**Preliminary Breath Test Device  
Monthly Verification Log**

Department NYPD

Device Serial # 052737

Test Date	Sim. Solution Lot #	Simulator Temperature			Sim. Serial #	Test Results			Officer Performing Verification	Sim. Solution Change *
		1	2	3		1	2	3		

2/7/20	19750	33.9	33.9	33.9	3701	.09	.09	.09	Kearns	—

\* - Please enter a √ to denote a simulator solution lot number change