

SOLUTION CHANGE

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 04/29/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:39
INTERNAL STANDARD	VERIFIED	18:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:40
BLANK TEST	0.00	18:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:41
BLANK TEST	0.00	18:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:42
BLANK TEST	0.00	18:43

