

SUPERVISOR TEST

NEW YORK MILLS  
POLICE DEPT.  
DATAMASTER dmt: 500475

Date: 06/24/2020

OPERATOR NAME:  
ROBERT C FRANKLAND  
Permit #: 40674  
Operator Agency/Dept: NY MILLS PD  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:58
INTERNAL STANDARD	VERIFIED	21:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:58
BLANK TEST	0.00	21:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:59
BLANK TEST	0.00	22:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:01
BLANK TEST	0.00	22:02

