

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197




WFO

Date: 03/04/2020

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:54
INTERNAL STANDARD	VERIFIED	20:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:55
BLANK TEST	0.00	20:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:56
BLANK TEST	0.00	20:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:57
BLANK TEST	0.00	20:58



SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197




Date: 03/11/2020

OPERATOR NAME:
MICHAEL P GIBBONS
Permit #: 45972
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:38
INTERNAL STANDARD	VERIFIED	00:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:38
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:40
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:41
BLANK TEST	0.00	00:42

*Datamaster
WFO Weekly
Test*



SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

LATIC

Date: 03/12/2020

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:20
INTERNAL STANDARD	VERIFIED	03:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:20
BLANK TEST	0.00	03:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:22
BLANK TEST	0.00	03:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:23
BLANK TEST	0.00	03:24

