

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197




Date: 03/19/2020

OPERATOR NAME:
MICHAEL P GIBBONS
Permit #: 45972
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:58
INTERNAL STANDARD	VERIFIED	21:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:58
BLANK TEST	0.00	21:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:00
BLANK TEST	0.00	22:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:01
BLANK TEST	0.00	22:02

WFO
Data master
weekly test



SUPERVISOR TEST

WFO

Cornish

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/25/2020

OPERATOR NAME:
ROBERT S CORNISH
Permit #: 48068
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:43
INTERNAL STANDARD	VERIFIED	23:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:45
BLANK TEST	0.00	23:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:46
BLANK TEST	0.00	23:47

