

SUPERVISOR TEST

BFO



ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/10/2020

OPERATOR NAME:
BRETT A GRABELDINGER
Permit #: 46896
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:54
INTERNAL STANDARD	VERIFIED	22:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:55
BLANK TEST	0.00	22:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:56
BLANK TEST	0.00	22:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:57
BLANK TEST	0.00	22:58



BFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 06/17/2020

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:37
INTERNAL STANDARD	VERIFIED	23:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:37
BLANK TEST	0.00	23:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:38
BLANK TEST	0.00	23:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:40
BLANK TEST	0.00	23:40

