

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LATIC

Date: 03/25/2020

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:26
INTERNAL STANDARD	VERIFIED	18:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:27
BLANK TEST	0.00	18:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:28
BLANK TEST	0.00	18:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:29
BLANK TEST	0.00	18:30

