

SUPERVISOR TEST

LATIC

CPD

ONEIDA CO.  
SHERIFF'S DEPT  
DATAMASTER dmt: 123906

Date: 01/30/2020

OPERATOR NAME:  
EMRAH LATIC  
Permit #: 46908  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:01
INTERNAL STANDARD	VERIFIED	02:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:02
BLANK TEST	0.00	02:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:03
BLANK TEST	0.00	02:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:04
BLANK TEST	0.00	02:05

