

SUPERVISOR TEST

SHERRILL CITY
POLICE DEPT
DATAMASTER dmt: 115108

Date: 02/12/2020

OPERATOR NAME:
ROBERT J DRAKE
Permit #: 27298
Operator Agency/Dept: SHERRILL CITY PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:30
INTERNAL STANDARD	VERIFIED	10:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:31
BLANK TEST	0.00	10:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:32
BLANK TEST	0.00	10:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:33
BLANK TEST	0.00	10:34

