

SUPERVISOR TEST

SHERRILL CITY
POLICE DEPT
DATAMASTER dmt: 115108

Date: 02/19/2020

OPERATOR NAME:
ROBERT J DRAKE
Permit #: 27298
Operator Agency/Dept: SHERRILL CITY PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:16
INTERNAL STANDARD	VERIFIED	07:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:16
BLANK TEST	0.00	07:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:18
BLANK TEST	0.00	07:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:19
BLANK TEST	0.00	07:20

