

SUPERVISOR TEST

-----  
SHERRILL CITY  
POLICE DEPT  
DATAMASTER dmt: 115108  
-----

Date: 05/20/2020

OPERATOR NAME:  
ROBERT J DRAKE  
Permit #: 27298  
Operator Agency/Dept: SHERRILL CITY PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:27
INTERNAL STANDARD	VERIFIED	08:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:27
BLANK TEST	0.00	08:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:29
BLANK TEST	0.00	08:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:30
BLANK TEST	0.00	08:31

