

SOLUTION CHANGE

NEW HARTFORD TOWN
POLICE DEPT
DATAMASTER dmt: 127606

Date: 01/28/2020

OPERATOR NAME:
MICHAEL T KOWALSKI
Permit #: 31077
Operator Agency/Dept: NEW HARTFORD TOWN PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:13
INTERNAL STANDARD	VERIFIED	14:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:13
BLANK TEST	0.00	14:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:15
BLANK TEST	0.00	14:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	14:16
BLANK TEST	0.00	14:17

