

SOLUTION CHANGE

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 04/01/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:04
INTERNAL STANDARD	VERIFIED	17:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:04
BLANK TEST	0.00	17:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:06
BLANK TEST	0.00	17:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:07
BLANK TEST	0.00	17:08

