

SOLUTION CHANGE

NEW YORK MILLS  
POLICE DEPT.  
DATAMASTER dmt: 500475

Date: 10/31/2019

OPERATOR NAME:  
ROBERT C FRANKLAND  
Permit #: 40674  
Operator Agency/Dept: NY MILLS PD  
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:19
INTERNAL STANDARD	VERIFIED	15:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:19
BLANK TEST	0.00	15:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:20
BLANK TEST	0.00	15:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:22
BLANK TEST	0.00	15:23

