

SUPERVISOR TEST

NEW YORK MILLS  
POLICE DEPT.  
DATAMASTER dmt: 500475

Date: 02/12/2020

OPERATOR NAME:  
ROBERT C FRANKLAND  
Permit #: 40674  
Operator Agency/Dept: NY MILLS PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:58
INTERNAL STANDARD	VERIFIED	18:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:59
BLANK TEST	0.00	19:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:00
BLANK TEST	0.00	19:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:01
BLANK TEST	0.00	19:02

