

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 12/25/2019

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:03
INTERNAL STANDARD	VERIFIED	18:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:04
BLANK TEST	0.00	18:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:05
BLANK TEST	0.00	18:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:06
BLANK TEST	0.00	18:07

